

difference absorption spectroscopy

A highly concentrated analyte in an analytical sample can be determined with better precision by replacing the *blank (reference) cell* by one containing a solution of the analyte or other absorber of known concentration; this is known as difference absorption spectroscopy. Difference spectra can also be obtained by computer or other subtraction methods.

1988, 60, 1455