

precision (of a balance)

The *standard deviation* of the instrument for the stated load (e.g. l_{20} for a 20 g load). A statement of the procedure, conditions, and experience of the observer should be included.

O.B. 36